



Request for Proposals 2023-028

First Aid and Service of AED's

ADDENDUM No. 1

ISSUE DATE: December 12, 2022

Responding offerors on this project are hereby notified that this Addendum shall be made a part of the above named RFP document.

The following items add to, modify, and/or clarify the RFP documents and shall have the full force and effect of the original Documents. This Addendum shall be acknowledged by the Offeror in the RFP document.

A. Add/Delete/Replace

1. **Add to 4.2 First Aid Cabinets standard stock; page 12 the following:**
"Medical examiner gloves (2 pair)"

B. Question/Answer Section

1. **Question:** Can you be more specific on what this item is?
 - Glucose
 - Eye Buffer Solution
 - Junior Strips
 - Tourniquet, Splint - universal or specific?

Answer:

- Glucose - Glucose, which elevates blood sugar for Diabetes.
- Eye Buffer Solution- A type of eyewash that can be used in case someone gets something in their eye
- Junior Strips- Smaller band aids
- Tourniquet, Splint - universal or specific? - Universal

2. **Question:** Quantity per First Aid Kits? 1 per box?

Answer: The description of what we're requesting is listed in section 4.2 of the RFP .

3. **Question:** Can you please supply a current list of all the AED's pad expiration dates?

Answer: Refer to attachment No. 2. This is our current list.


4. **Question:** Could you please confirm if the AED's are Lifeline or Lifeline Views?

Answer: They are Defibtech Lifeline AED's

Attachment No.1: Sign-in Sheet - Pre-Bid Site Visit

Attachment No.2: AED Pad Expiration Dates

End of Addendum No. 1



Pre-Proposal Conference
RFP 2023-028 First Aid and Service of AED's
Location: Union County Government Center, 500 N Main St, Monroe, NC
Monday, November 28, 2022, 9:00 AM

Proposal Due Date & Time: December 15, 2022 at 2:00 PM EDT
Electronic Submission

Name	Company	Email	Phone
Nathan Murphy	Cintas	murphy@cintas.com	704-391-2718
Angela Wernmurry	Thompson Safety	awernmurry@thompson-safety.com	704-517-5809
Chris Padgett	UniFirst First Aid + Safety	christopher-padgett@unifirst.com	704-230-8273
Gabrielle Bethel	UniFirst First Aid + Safety	Gabrielle_Bethel@unifirst.com	704-499-0814

**UNION COUNTY, NORTH CAROLINA
AED'S**

Attachment No.2: AED Pad Expiration Dates

ADDRESS	LOCATION	UNIT TYPE	DATE UNIT INSTALLED	PAD 1	PAD 2	BATTERY	CHILD PADS
AGRICULTURAL CENTER 3230 PRESSON ROAD MONROE, NC 28112	Admin Area	Reviver	7/17	3/24	3/24	5/27	
	Banquet Hall	Reviver	7/17	3/24		2/25	
	Prefunction Area	Reviver	7/17	3/24	3/24	2/25	
ANIMAL SHELTER 3340 PRESSON ROAD MONROE, NC 28112		Reviver		2/24		8/27	4/24
	Adoption	Reviver	7/17	1/24		10/27	
BOARD OF ELECTIONS 316 E WINDSOR ST. MONROE, NC 28112	Old Section	Reviver	7/17	2/24		2/25	
	New Addition		7/22	7/24		7/27	
CANE CREEK PARK 5213 HARKEY RD WAXHAW, NC 28173	Office	Reviver		1/24		4/24	
	Putt Hutt	Reviver	2/19	2/23		4/24	
	Camp Store	Reviver	4/19	6/23	7/23	9/24	
FARMERS MARKET 805 SKYWAY DR. MONROE, NC 28110	Office 1st floor	Reviver	7/17	1/24		10/27	
	Hall 1st floor	Reviver	2/21	2/23		2/26	
GOVERNMENT CENTER 500 N MAIN ST. MONROE, NC 28112	Ground Floor	Reviver	7/17	2/24		10/25	
	911 Call Center	Reviver	8/15	3/24		1/26	
	1st Floor	Reviver	6/07	2/24	2/24	6/27	
	2nd Floor	Reviver	2/08	2/24	2/24	2/24	
	3rd Floor	Reviver	7/17	2/24		10/27	
	4th Floor	Reviver	6/07	3/24	1/23	2/26	
	5th Floor	Reviver	7/17	2/24		10/27	
	6th Floor	Reviver	2/09	2/24		2/25	
	7th Floor	Reviver	7/17	2/24		6/27	
8th Floor	Reviver	7/17	2/24		1/27		
9th Floor	Reviver	5/17	3/24		2/25		

**UNION COUNTY, NORTH CAROLINA
AED'S**

Attachment No.2: AED Pad Expiration Dates

ADDRESS	LOCATION	UNIT TYPE	DATE UNIT INSTALLED	PAD 1	PAD 2	BATTERY	CHILD PADS
GROUP HOME 600 HONEYSUCKLE LN. MONROE, NC 28110		Reviver	7/17	1/24		10/27	
HISTORIC COURTHOUSE 300 N MAIN ST. MONROE, NC 28112		Reviver	7/17	1/24		6/27	
					8/23		
JUDICIAL CENTER 400 NORTH MAIN ST. MONROE, NC 28112	Intake	Reviver	7/16	2/24		1/27	
	1st Floor	Reviver	10/04	2/24	2/24	6/27	
	2nd Floor	Reviver	7/16	2/24	2/24	2/27	
	3rd Floor	Reviver	10/04	2/24	2/24	6/27	
	4th Floor	Reviver	5/17	2/24		5/27	
	4th Floor DA's Office	Reviver	12/13	1/23		3/24	
WAXHAW LIBRARY 509 S PROVIDENCE ST. WAXHAW, NC 28173		Reviver	7/17	2/24		4/24	2/24
HISTORIC POST OFFICE 407 N MAIN ST. MONROE, NC 28112	Veterans Services Office	Reviver	6/12	7/23		6/27	
UNION WEST LIBRARY 123 UNIONVILLE-INDIAN TRAIL RD. INDIAN TRAIL, NC 28079		Reviver	7/07	3/24		4/27	1/23
MAIN LIBRARY 316 E WINDSOR ST. MONROE, NC MONROE, NC 28112		Reviver	7/17	3/24	2/23	8/26	
LOIS MORGAN EDWARDS LIBRARY 414 HASTY ST. MARSHVILLE, NC 28103		Reviver		2/24		10/27	2/24

**UNION COUNTY, NORTH CAROLINA
AED'S**

Attachment No.2: AED Pad Expiration Dates

ADDRESS	LOCATION	UNIT TYPE	DATE UNIT INSTALLED	PAD 1	PAD 2	BATTERY	CHILD PADS

**UNION COUNTY, NORTH CAROLINA
AED'S**

Attachment No.2: AED Pad Expiration Dates

ADDRESS	LOCATION	UNIT TYPE	DATE UNIT INSTALLED	PAD 1	PAD 2	BATTERY	CHILD PADS
PUBLIC WORKS LANDFILL 2125 AUSTIN CHANEY RD WINGATE, NC 28174		Reviver		2/24		2/25	
PUBLIC WORKS OPERATIONS CTR 4600 GOLDMINE ROAD MONROE, NC 28110	Admin Bldg	Reviver		2/23		2/25	
	Warehouse	Reviver		1/23		2/25	
SHERIFF'S OFFICE 3344 PRESSON RD. MONROE, NC 28112	Master Command	Reviver		1/23	3/24	5/26	
	Nurse	Reviver		3/24	3/24	5/26	
PATTON AVENUE COMPLEX 610A PATTON AVE. MONROE, NC 28110	Transportation Dept	Reviver		2/23		6/27	
JESSE HELMS PARK 3224 PRESSON RD. MONROE, NC 28112			3/19	6/23		4/24	
HUMAN SERVICES BUILDING 2330 CONCORD AVENUE MONROE, NC	D302 (Break Room)	Reviver	12/17	3/24		9/26	
	A202 (Front Desk)	Reviver	12/17	3/24	11/24	12/25	7/24
	B231 (Break Room)	Reviver	12/17	3/24		3/23	
	D202 (Break Room)	Reviver	12/17	3/24		3/23	
	E101 (Waiting Room)	Reviver		3/24		3/23	
	Nurses Station (crash cart)	Reviver	3/18	3/24		5/23	2/23
	Nurse Station (extra)	Reviver		9/23		6/23	4/23
	Nurse (women's health)	Reviver		4/24		6/27	4/23
Nurse (women's health)	Reviver		4/24		9/27		
EVENT CENTER AT JESSE HELMS PARK 307 CULTIVATION CIRCLE MONROE, NC 28112	Office 119		5/21	5/23		5/26	

**UNION COUNTY, NORTH CAROLINA
AED'S**

Attachment No.2: AED Pad Expiration Dates

ADDRESS	LOCATION	UNIT TYPE	DATE UNIT INSTALLED	PAD 1	PAD 2	BATTERY	CHILD PADS
PROGRESS BUILDING 1407 AIRPORT RD MONROE, NC 28110	Break Room 107 Main Warehouse		3/21 3/21	4/23 4/23		4/26 4/26	

ITEM

PADS

BATTERIES

DEFIBTECH'S

REPLACEMENT SCHEDULE

2 YEARS

5 YEARS

8 YEARS WARRANTY ENDS (10-15 YEAR REPLACEMENT)